

PUBLIC HEALTH UNIT INFECTION PREVENTION AND CONTROL LAPSE REPORT

Initial Report

Premise/facility under investigation (name and address)	Greengate Medical Centre 170 Silvercreek Parkway N, unit 2 Guelph ON, N1H 7P7
Type of premise/facility: (E.g. clinic, personal services setting)	Family Practice
Date Board of Health became aware of IPAC lapse	December 14, 2018
Date of Initial Report posting	
Date of Initial Report update(s) (if applicable)	
How the IPAC lapse was identified	A complaint made by a member of the public
Summary Description of the IPAC Lapse	During the complaint inspection it was observed that the cleaning, disinfection and sterilization of medical devices on site did not follow Provincial Infectious Disease Advisory Committee (PIDAC) Best Practice Standards for medical device reprocessing and infection control in a clinical office setting.
IPAC Lapse Investigation	
Did the IPAC lapse involve a member of a regulatory college?	Yes
If yes, was the issue referred to the regulatory college?	Reports were made to the College of Physicians and Surgeons of Ontario and Ontario College of Pharmacists
Were any corrective measures recommended and/or implemented?	The family practice is to ensure: <ul style="list-style-type: none"> - Medical device reprocessing and infection control practices will follow the IPAC best practice documents provided by the Provincial Infectious Diseases Advisory Committee (PIDAC) and Canadian Safety Association (CSA). - Internal policies and procedures on infection control and medical device reprocessing are in place and followed. - Expired products are discarded

Please provide further details/steps	- Staff responded immediately to implement IPAC practices as per current Provincial Infectious Disease Advisory Committee (PIDAC) guidelines for cleaning and reprocessing of medical equipment.
Date any order(s) or directive(s) were issued to the owners/operators (if applicable)	N/A
Initial Report Comments and Contact Information	
Any Additional Comments (Do not include any personal information or personal health information)	
If you have any further questions, please contact:	
Name	Janice Walters
Title	Program Manager
E-mail address	janice.walters@wdgpublichealth.ca
Phone number	519-822-2715 ext. 4246
Final Report	
Date of Final Report posting:	January 3, 2019
Date any order(s) or directive(s) were issued to the owner/operator (if applicable)	N/A
Brief description of corrective measures taken	On January 3, 2019 a re-inspection was conducted. At this time IPAC practices were in compliance with IPAC Best Practices. Patient care services using reusable equipment was allowed to resume.
Date all corrective measures were confirmed to have been completed	January 3, 2019
Final Report Comments and Contact Information	
Any Additional Comments (Do not include any personal information or personal health information)	
If you have any further questions, please contact:	
Name	Janice Walters

Title	Program Manager
Email address	janice.walters@wdgpublichealth.ca
Phone number	519-822-2715 ext. 4246